
**Performance**

Measuring range :	0.1 to 10 ppm
Sampling time :	1 to 10 hours
Colour change :	Pink → Yellow
Corrections for temperature & humidity :	Temperature & humidity corrections are necessary.
Relative standard deviation :	10 % (for 1 to 10 ppm·hr)
Shelf life :	2 years

Reaction principle**Possible coexisting substances and their interferences** (NOTE : Page 2-5)

Substance	Concentration	Interference	Changes colour by itself to
Amines, Hydrazine		+	Yellow

Other substances measurable with this Dosi-tube

Substance	Correction	Sampling time	Measuring range
Methylamine	Factor : 1.9	} 1 to 10 hours	0.19 to 19 ppm
Trimethylamine	Factor : 2.3		0.23 to 23 ppm

Calibration gas generation

Permeation tube method