

**Performance**

Measuring range	0.5 to 30 ppm
Number of pump strokes	1 (100 ml)
Correction factor	1
Sampling time	2 min
Detecting limit :	0.1 ppm (1 pump stroke)
Colour change :	White → Yellowish orange
Corrections for temperature & humidity :	Unnecessary
Shelf life :	3 years

Reaction principle

Pyrotec : Nitro compounds (Pyrolyzing) NO_x

Pyrotube : NO_x + CrO₃ + H₂SO₄ → NO₂

NO₂ + o-Tolidine → Yellowish orange product

Possible coexisting substances and their interferences (NOTE : Page 2-5)

Substance	Concentration	Interference	Changes colour by itself to
Chlorine dioxide		+	} Yellowish orange
Halogens		+	
Halogenated hydrocarbons		+	
Hydrogen chloride		+	
Sulphur dioxide	≥ 25 ppm	– (Bleaching)	No
Hydrogen sulphide	≥ 25 ppm	– (Bleaching)	No

Substances measurable with this Pyrotube

Substance	n	Correction factor	Measuring range
Acetonitrile	1	6.0	3 to 180 ppm
Nitroethane	1	8.0	4 to 240 ppm
Nitrogen dioxide	1	1.0	0.5 to 30 ppm
Nitromethane	1	10.0	5 to 300 ppm
1-Nitropropane	1	8.4	4.2 to 252 ppm
2-Nitropropane	1	7.4	3.7 to 222 ppm

Calibration gas generation

Permeation tube method